


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-5B-C-N*

S.N.: *4164*

Date: *FEB 11 1988*



- a) Output signal amplitude:
0 TO 48 AMP TO
- b) Pulse width: *0.75 μ s*
25 TO 500 NS
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

$R_L = 0.7 \Omega$
40 db ATTEN
 $\therefore \approx 10$ VOLTS/DIV

120/240V
50-60 Hz
[Signature]