


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR-3-PW-C-P

S.N.: 4169

Date: JAN 28 1988



Ⓐ  $R_L = 50 \Omega$   
PRF = 1 kHz



Ⓑ  $R_L = 50 \Omega$   
PRF = 50 kHz

- a) Output signal amplitude:  
0 TO +200 V
- b) Pulse width:  
0.1 TO 100  $\mu$ SEC
- c) Rise time:  
 $\leq 10$  NSEC
- d) Fall time:  
 $\leq 10$  NSEC
- e) PRF:  
MIN PW TO 1.0  $\mu$ SEC; 0 TO 1 kHz  
MAX PW; 50 Hz.
- f) Jitter, stability:  
OK
- g) Prime power:  
120/250 V  
50-60 Hz