


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A-1-N*

S.N.: *4172*

Date: *MAR 1 1988*



- a) Output signal amplitude:
0 TO -200 VOLTS.
- b) Pulse width:
50 NSEC TO 1.0 USEC
(0.5% MAX DUTY CYCLE)
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
MIN PW: 0 TO 100 KHz
MAX PW: 0 TO 5 KHz
- f) Jitter, stability: *(0.5% MAX DUTY CYCLE)*
OK
- g) Prime power:

R_L = 50 Ω
1.0 KHz

+24 ± 1.0 VOLTS.
500 mA (MAX).