


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-N*

S.N.: *4175*

Date: *FEB 23 1978*



- d) Output signal amplitude:
-25 to -100 mV
- e) Pulse width:
≤ 25 ns
- f) Rise time:
NA
- g) Fall time:
NA
- h) PRF:
0 to 1 kHz
- i) Jitter, stability:
OK
- j) Prime power:

*≈ 60 mV/div
500 kHz*

*+15 ± 1.0 volts
250 mA (max)*

[Signature]