


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AUM-3-PS-N-D-M*

S.N.: *4176*

Date: *FEB 24 1988*



- a) Output signal amplitude:
0 TO -15 VOLTS
- b) Pulse width:
2 TO 15 NS
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*40 dB ATTEN.
10 VOLTS DIV
= 12 MHz*

*120 / 240 V
50-60 Hz
A*