


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVM4.3-P

S.N.: 4179

Date: MAR 10 1988



- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width:
≤ 600 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

50 lb APPROX
≈ 32 VOLTS/DIV
700 MHz.

+24 ± 1.0 VOLTS
500 mA (MAX)
[Signature]