


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *APR-3-PW-C-P*

S.N.: *4180*

Date: *JAN 29 1988*



Ⓐ  $R_L = 50 \Omega$   
 $1 \text{ KHz}$



Ⓑ  $R_L = 50 \Omega$   
 $50 \text{ Hz}$

- a) Output signal amplitude:  
*0.70 + 200 V.O.V.B*
- b) Pulse width:  
*0.1 TO 100 nSEC.*
- c) Rise time:  
*≤ 10 nSEC*
- d) Fall time:  
*≤ 10 nSEC*
- e) PRF:  
*0.1 TO 1.0 USEC PW: 0 TO 1KHz*  
*MAX PW: 50 Hz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*

*[Signature]*