



PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-C-PN-OS-M
S.N.: 4124
Date: MAR 2 1988



Ⓐ 50 db ATTEN
- 32 VOLTS/DIV
20 KHz



Ⓑ AS Ⓐ BUT 1.0 nSEC/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
2 TO 100 nSEC
- c) Rise time:
 ≤ 350 pSEC
- d) Fall time:
 ≤ 2 nSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.

[Signature]