



PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-V-P*
S.N.: *4186*
Date: *MAR 3 1 00 00*



*(A) 40 dB ATTN
10 VOLTS/DIV
50 KHz*

- a) Output signal amplitude:
7.5 VOLT
- b) Pulse width:
2 TO 50 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 50 KHz



- f) Jitter, stability:
OK
- g) Prime power:
*7.5 ± 1.0 VOLT
300 mA (MAX)*

*(B) RISE TIME
10 VOLTS/DIV
50 KHz*