


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PAH-143-P*

S.N.: *4192*

Date: *MAR 8 1988*



*50 dB ATTEN  
≈ 32 VOLTS/DIV  
500 kHz*

- a) Output signal amplitude:  
*0 TO +60 VOLTS*
- b) Pulse width:  
*≤ 1.0 nSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 ± 1.0 VOLT  
200 mA (MAX)*