


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-D-LIA*
S.N.: *4197*
Date: *MAR 17 88*



- a) Output signal amplitude:
0 TO +50.00V
- b) Pulse width:
5 TO 50 NSEC.
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 ± 1.0 00V.

50 db ATTEN
32 VOLTS/DIV.
50 KHz.

200 mA (MAX!)
[Signature]