


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APR-3-pw-c-p*

S.N.: *4205*

Date: *FEB 24 1988*



Ⓐ $R_L = 50 \Omega$
 50 kHz



Ⓑ $R_L = 50 \Omega$
 100 kHz

- a) Output signal amplitude:
0 to +200 VOLTS
- b) Pulse width:
0.1 TO 100 μ SEC
- c) Rise time:
 $\leq 10 \text{ NSEC}$
- d) Fall time:
 $\leq 10 \text{ NSEC}$
- e) PRF:
MIN PW: 0 TO 1 kHz
MAX PW: 0 TO 50 kHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 kHz