


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ACH-HV3-P*

S.N.: *4212*

Date: *MAR 8 1988*



- a) Output signal amplitude:
0 to +60 VOLTS
- b) Pulse width:
≤ 1.0 nSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*50 dB ATTEN
≈ 32 VOLTS/DIV
500 kHz*

*+15 ± 1.0 VOLT
200 mA (MAX)*

[Signature]