


PULSE GENERATOR

PERFORMANCE CHECK


Model: PAH-M1-P

S.N.: 4219

Date: FEB 11 1988



Ⓐ 50 dB ATTEN
32 VOLTS/DIV
500 KHz



Ⓑ 40 dB ATTEN
10 VOLTS/DIV
500 KHz

a) Output signal amplitude:
≥ +40 VOLTS*

b) Pulse width:

c) Rise time:
≤ 1 NSEC

d) Fall time:
NA

e) PRF:
NA

f) Jitter, stability:
0 TO 1 MHz

g) Prime power:
OK

+135 ± 5 VOLTS
5 mA (MAX)

* $V_{OUT} = 44$ VOLTS