


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVD-1-P*

S.N.: *4232*

Date: *FEB 23 '88*



*≈ 60 Amps/div.
500 Hz.*

- a) Output signal amplitude:
+25 TO +100 Amps
- b) Pulse width:
≤ 25 nsec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:

*+15 ± 1.0 VOLT.
250 mA (MAX)*

[Signature]