


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*

S.N.: *4233*

Date: *FEB 23 1988*



- a) Output signal amplitude:
72.5 mV ± 100 mV
- b) Pulse width:
6.25 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0.70 / kHz
- f) Jitter, stability:
OK
- g) Prime power:

*≈ 60 mW P_{AV}
300 Hz*

*+15 ± 1.0 VOLTS
250 mA (MAX)*

[Signature]