


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *A20-1-P*  
S.N.: *4234*  
Date: *FEB 23 1958*



- a) Output signal amplitude:  
*+25 TO +10.0 AMPS*
- b) Pulse width:  
*± 25 NSER.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 60 AMPS/DIV*  
*500 Hz.*

*+15 ± 1.0 VOLT.*  
*250 mA (max)*  
*[Signature]*