



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-AMP-P*
S.N.: *4237*
Date: *FEB 26 88*



Ⓐ *50 db ATTN*
32 VOLTS/DIV
MIN PW
1.0 MHz



Ⓑ *50 db ATTN*
32 VOLTS/DIV
50 KHz

- a) Output signal amplitude:
7.40 VOLTS
- b) Pulse width:
2 TO 100 NSER
- c) Rise time:
≤ 350 PSEZ
- d) Fall time:
≤ 2 NSER
- e) PRF:
MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
15 ± 1.0 VOLT
200 mA (MAX)

[Signature]