


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVC-A-1-PW-P-PS-GMI*
S.N.: *4244*
Date: *MAR 1 1988*



- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width: ($R_L \geq 1K$)
50 NSEC TO 1.0 USEC
- c) Rise time:
 ≤ 5 NSEC
- d) Fall time:
 ≤ 5 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

$R_L = 1K$
PRF = 5 KHz