



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-1-C-P*
S.N.: *4245*
Date: *APRIL 8 1988*



Ⓐ 60 db ATTN
≈ 100 VOLTS IN
≈ 10 KHz
MID VOLT



Ⓑ AS Ⓐ BUT
≈ 50 KHz & OUTPUT
MID RANGE

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
≤ 4 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]