



PULSE GENERATOR  
PERFORMANCE CHECK

Model: APMP-1-P-CA-EW-ED-TRF-ECL-M1  
 S.N.: 4247  
 Date: APRIL 12 1968



Ⓐ 40 db ATTEN  
 5 VOLTS  
 700 MHz  
 TR TR HOLD



Ⓑ AS Ⓐ BUT  
 TR, TR LOW

- a) Output signal amplitude:  
0 TO +5 VOLT.
- b) Pulse width:  
5 TO 50 NS (DUTY CYCLE MAX 20%)
- c) Rise time:  
0.1 OR 1.0 NS
- d) Fall time:  
0.1 OR 1.0 NS
- e) PRF:  
0 TO 10 MHz (MAX DUTY CYCLE 2%)
- f) Jitter, stability:  
OK
- g) Prime power:  
2.4 ± 1.0 VOLT.  
300 mA (MAX)

h) OFFSET 0 TO ±5 VOLT  
 TO 50 Ω