


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVL-2-F*

S.N.: *4249*

Date: *APRIL 7 1988*



- d) Output signal amplitude:  
*0 TO +350 V*
- e) Pulse width:  
*5 TO 100 nSEC*
- f) Rise time:  
*≤ 2 nSEC*
- g) Fall time:  
*≤ 3 nSEC*
- h) PRF:  
*0 TO 5 kHz*
- i) Jitter, stability:  
*OK*
- j) Prime power:

*70 db ATTEN*  
*≈ 320 VOLTS/DIV*  
*2 kHz*

*+15 ± 1.0 VOLT*  
*500 mA (MAX)*