


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APR-B2-C-PN-W-05-M1-M*

S.N.: *4253*

Date: *APRIL 21 88*



*P_{out} R_i = 50Ω
5 KHz*



*P_{out} R_i = 50Ω
5 KHz*

- a) Output signal amplitude:
0 TO ±50 VOLTS
- b) Pulse width:
0.1 TO 10 μsec
- c) Rise time:
≤ 10 nsec
- d) Fall time:
≤ 10 nsec
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*

[Signature]