


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVN-1-C-PN-M-OT*

S.N.: *4257*

Date: *APRIL 26 '88*



- a) Output signal amplitude:  
*0 TO ±5 VOLTS*
- b) Pulse width:  
*0.2 TO 1.0 NS*
- c) Rise time:  
*≤ 100 PSEZ*
- d) Fall time:  
*≤ 100 PSEZ*
- e) PRF:  
*50 TO 250 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 VOLTS*  
*50-60 MHz*

*40dB ATTEN*  
*5 VOLTS/DIV*  
*200 MHz*  
*PRF*

*[Signature]*