


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVA-S1-C-N

S.N.: 4262

Date: APRIL 27 88



- a) Output signal amplitude:
0 TO -10 V.
- b) Pulse width:
≤ 130 psec
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN

10 VOLTS/DIV

1.0 MHz PRF.

120/240 V

50-60 Hz

[Signature]