


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-1-C-MSU1*

S.N.: *4263*

Date: *MAY 3 88*



- a) Output signal amplitude:
0 TO +200V
($R_L \geq 500 \Omega$)
- b) Pulse width:
0.05 TO 2.0 μ SEC
- c) Rise time:
 ≤ 5 NSEC
- d) Fall time:
 ≤ 15 NSEC
- e) PRF:
0 TO 10 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

$R_L = 450 \Omega$
PRF = 10 KHZ

120/240 VOLTS
50-60 HZ
[Signature]