


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-HV3-P*

S.N.: *4265*

Date: *MAR 8 1988*



- d) Output signal amplitude:  
*0 TO +60.004V*
- e) Pulse width:  
*≤ 1.0 nSEC*
- f) Rise time:  
*NA*
- g) Fall time:  
*NA*
- h) PRF:  
*0 TO 1.0 MHz*
- i) Jitter, stability:  
*OK*
- j) Prime power:  
*+15 ± 1.0 uA.*

*50 dB atten*

*≈ 32 VOLTS/DIV*

*500 kHz*

*+15 ± 1.0 uA.*

*200 mA (max)*