


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AMP-1A-PS-P-W-6GB1*

S.N.: *4280*


Date: *FEB 29 88*



Ⓐ *40 db ATTEN
5 VOLTS/DIV
50 KHZ*




Ⓑ AS Ⓐ BUT RISE TIME



Ⓒ AS Ⓐ BUT FALL TIME

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
*< 100 nSEC TO 10.00 μSEC
(10% MAX DUTY CYCLE)*
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 150 pSEC
- e) PRF: *0 TO 100 KHZ.
(10% MAX DUTY CYCLE)*
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*



Ⓓ REAL TIME SCOP
DISPLAY.