

PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-M1-P

S.N.: 4282

Date: APRIL 6 '88



- a) Output signal amplitude:
 $\geq +40$ VOLTS
- b) Pulse width:
 ≤ 1 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN

10 VOLTS/DIV

≈ 59 VOLTS

500 KHz

+135 \pm 5 VOLTS
(5 mA MAX)