


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-U-HV3-CP-M-CA-EW*

S.N.: *4283*

Date: *MAY 27 1988*



Ⓐ 60 db ATTEN
100 VOLTS/DIV
2.5 KHZ



Ⓑ AS Ⓐ BUT 1NS/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 200 PSEZ
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ