


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-P.*

S.N.: *4290*

Date: *APRIL 5 1988*



*≈ 16 amps / DIV
500 Hz.*

- a) Output signal amplitude:
10 TO 50 AMPS
- b) Pulse width:
≈ 20 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*+15 ± 1.0 VOLTS
2.50 mA (max)*