

PULSE GENERATOR  
PERFORMANCE CHECK

Model:

AMP-1-PW-C-17

S.N.:

4294

Date:

MAY 3 1948

a) Output signal amplitude:

0 TO 720 VOLTS

b) Pulse width:

0.1 TO 5.0 USEC

c) Rise time:

≤ 10 NSEC

d) Fall time:

≤ 10 NSEC

e) PRF:

0 TO 1 KHZ

f) Jitter, stability:

OK

g) Prime power:

120 / 240 V

50-60 Hz

$R_c = 50 \Omega$

1 KHZ PRF.