


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PLM-1-P*
S.N.: *4307*
Date: *MAY 13 1988*



- a) Output signal amplitude:
0 TO 4.5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 PSEZ
- d) Fall time:
≤ 135 PSEZ
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40db ATTEN
5 VOLTS/DIV
7.0 MHz

+24 ± 1.0 VOLTS
500 mA (MAX)

[Signature]