


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVNN-1-C-PN-M-OT*


S.N.: *4300*

Date: *APR 9 88*



Ⓐ

*40 dB ATTEN
5 VOLTS/DIV
≈ 120 MHz*



Ⓑ

*AS Ⓐ BUT
1 nSEC/DIV*

- a) Output signal amplitude:
0 TO 5 VOLTS
- b) Pulse width:
*1-30 nSEC
(30-70% DUTY CYCLE)*
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 100 pSEC
- e) PRF:
25 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*