


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PV-107C-M-PW-N3*


S.N.: *4312*

Date: *MAY 17 1988*



- a) Output signal amplitude:
0 to ± 15 AMPS
- b) Pulse width:
50 nSR TO 1.0 μSR
- c) Rise time:
≤ 10 nSR
- d) Fall time:
≤ 10 nSR
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

Ⓐ *P_{out} R_L = 3 Ω*
± 3.3 AMPS/DIV
1.0 KHz



Ⓑ *N_{out} R_L = 3 Ω*
± 3.3 AMPS/DIV
1.0 KHz

[Signature]