



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MMMP-3-e-7*  
 S.N.: *4314*  
 Date: *MAY 13 1978*



⊙ 40 dB ATTEN  
 10 VOLTS/DIV  
 100 KHz



⊙ 1ns ⊙ BUT  
 1 nSEC/DIV  
 (RISE TIME)

- a) Output signal amplitude:  
*0 TO +20 VOLTS*
- b) Pulse width:  
*5 TO 100 nSEC*
- c) Rise time:  
*≤ 200 pSEC*
- d) Fall time:  
*≤ 300 pSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*