


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*
S.N.: *4316*
Date: *APRIL 5, 1988*



- a) Output signal amplitude:
+25 TO +100 AMP
- b) Pulse width:
~ 25 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

~ 60 AMPS / DIV
500 Hz.

+15 ± 1.0 VOLT
250 mA (MAX)