


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1-P*  
S.N.: *4317*  
Date: *APRIL 5 1988*



- a) Output signal amplitude:  
*+25 TO +100 mVPPS*
- b) Pulse width:  
*≈ 25 nS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 60 mVPPS / DIV*  
*500 Hz.*

*+15 ± 1.0 VOLT*  
*250 mA (MAX.)*  
*[Signature]*