


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*
S.N.: *4318*
Date: *APRIL 5 1988*



- a) Output signal amplitude:
+25 TO +100 AMP
- b) Pulse width:
≈ 25 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 K/Hz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 60 AMP/PS/DIV
500 Hz

+15 ± 1.0 004
250 mA (MAX)
[Signature]