


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PM-2-C-N*

S.N.: *4323*

Date: *APRIL 25 1988*



*70 db ATTEN*

*≈ 320 VOLTS PK*

*≈ 2 KHz PRF*

- a) Output signal amplitude:  
*0 TO -350 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*