


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVI-V-C-05-P*

S.N.: *4331*

Date: *MAY 13 1988*



Ⓐ 50 dB ATTEN  
0.32 VOLTS/DIV  
20 KHz



Ⓑ AS Ⓐ BUT  
1 nSEC/DIV  
(RISE TIME)

- a) Output signal amplitude:  
*0 TO +50 VOLTS*
- b) Pulse width:  
*3 TO 100 nSEC*
- c) Rise time:  
*≤ 350 pSEC*
- d) Fall time:  
*≤ 2 nSEC*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*A.*