


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVMP-3-C-CA-EW-ED-P*

S.N.: *4332*

Date: *MAY 20 1988*



② *40 db ATTEN*
10 VOLTS/DIV
1.0 ns/DIV



③ *AS ② BUT 1.0 ns/DIV*
(RISE TIME)

- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
8 TO 100 nSEC
- c) Rise time:
≤ 200 pSEC
- d) Fall time:
≤ 300 pSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]