


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-2-N-PS-DC1-TN-MHV*

S.N.: *4339*

Date: *MAY 31 1968*



- a) Output signal amplitude:
0 TO -350 VOLTS
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

70 db ATTEN
≈ 320 VOLTS/DIV
≈ 2.0 KHZ

120 / 240 V
50-60 HZ

h) tp: 110 NSEC.
(TN = 100 NSEC.
DC1 OPTION ADDS
10 NSEC)