


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVL-2A-C-T-W-P-M*

S.N.: *4340*

Date: *JUNE 2 1988*



*Ⓐ 60 db ATTEN  
∴ 100 VOLTS/DIV  
5 KHZ  
MIN TR*



*Ⓑ AS Ⓐ BUT  
MIN TR*

- a) Output signal amplitude:  
*0 TO +170 VOLTS*
- b) Pulse width:  
*0.4 TO 400 NSR*
- c) Rise time:  
*≤ 2 NSR*
- d) Fall time:  
*≤ 3 NSR*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V  
50-60 HZ*