


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVMM-3-C-PN-TR-CA-EW-NIM*

S.N.: *4345*


Date: *JUNE 20 88*



Ⓐ *OUT 40 dB ATTEN  
10 VOLTS/DIV  
12 MHz*



Ⓑ *TR out POSITION 1*



Ⓒ *AS Ⓑ BUT  
POSITION 5*

a) Output signal amplitude:

*0 TO ± 10 VOLTS*

b) Pulse width:

*1-5 NS*

c) Rise time:

*0.3 TO 2.0 NS*

d) Fall time:

*0.6 TO 2.0 NS*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+24 ± 10 VOLTS  
800 mA (MAX)*