


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-M2-C-M-OS-PN
S.N.: 4347
Date: JUNE 14 1988



60 db ATTEN
100 VOLTS/DIV
5. 1K Hz

- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
4 TO 100 NS
- c) Rise time:
≤ 500 PS
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 Hz

AA