


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-C-N-M-6W

S.N.: 4351

Date: MAY 26 1988



- a) Output signal amplitude:
0 TO -350 V.
- b) Pulse width:
5 TO 100 nSEC
- c) Rise time:
≤ 2 nSEC
- d) Fall time:
≤ 3 nSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

70 dB ATTEN
320 VOLTS/DIV
1.0 KHz

120/240 V
50-60 Hz

[Signature]