


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APD-2W-C-PN*

S.N.: *4354*

Date: *JUNE 2 1988*



- a) Output signal amplitude:
0 TO ± 10 AMPS
- b) Pulse width:
3 TO 50 nSEC
- c) Rise time:
5 / nSEC
- d) Fall time:
5 2 nSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

± 7 AMPS / DIV

20 KHz PRF

120/240 V

50 - 60 Hz