


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MM-3-C-PN*

S.N.: *4356*

Date: *JUNE 15 1988*



- a) Output signal amplitude:
0 TO ± 15 VOLTS
- b) Pulse width:
2 TO 15 NSEC.
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS/DIV
≈ 12 MHz PRF

120/240 V
50-60 Hz