


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-M1-P-40

S.N.: 4363

Date: JUNE 8, 1988



- a) Output signal amplitude:
 $\geq +40$ VOLTS
- b) Pulse width:
 ≤ 1.0 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHZ
- f) Jitter, stability:
OK
- g) Prime power:
 $+135 \pm 5.0$ VOLTS
5 mA (MAX)

40 db ATTEN

10 VOLTS/DIV

700 KHz PRF